

# Post Market Surveillance

**Mohammed Y Majrashi, PhD**  
**Executive Director, Surveillance**  
**and Biometrics**  
**Medical Devices Sector, Saudi FDA**

# Contents

- **Part A:**

**Introduction about Post-market Surveillance**

- **Part B:**

**Reactive system**

- **Adverse Events**
- **FSNs /FSCAs**
- **Safety communications**


- **Part C:**

**Risk Assessment and Clinical Evaluation Studies**

- **Part D:**

**Safe use of Medical Devices**

# Medical Devices Life Cycle



# (Post-market Surveillance)

Group of activities carried out by the regulatory authorities and manufacturers to get the necessary data to continuously assess the performance and safety of its medical devices after it has been marketing to ensure the compliance of these medical devices with the regulation requirements of medical devices.



# A pictorial view of PMS



Post-Market Surveillance  
Information is used for:

- Injury prevention
- Product improvement
- Development of standards
- Regulatory refinement

# Postmarket Surveillance Inputs

Int'l and local establishments  
(Manufacturers/AR ,  
importers and Distributors )

Healthcare facilities & public  
complains

**Input  
Sources**

Evaluation onsite visits (Safe  
use of Medical Devices)

Other int'l regulatory  
authorities


# Adverse Events

# What is MD Incident (Adverse Event)?

- **Adverse event:** means any malfunction or deterioration in the characteristics and/or performances of a medical device, including any inadequacy in its labeling or the instructions for use which may lead to compromise the health or safety of patients, users or third parties.
- **Reportable Adverse event:** any adverse event or any technical or medical reason leading to a Field Safety Corrective Action, which, directly or indirectly, might lead to or may have led:
  - (a) to the death of a patient, a user or another person or
  - (b) to a serious deterioration in their state of health.




# Reportable Adverse event



# Examples of Adverse Events


## Examples of Medical Devices Problems



- **Defects :**
  - IV pump bracket found with large crack and sharp edges
  - Surgical gloves found discolored and with holes




# Examples of Medical Devices Problems



- **Software problems**


- CT scanner found to have a software glitch in new version



- Virus infects device operating software




## Examples of Medical Devices Problems




- **Failure to work as intended /malfunction:**
  - Electrosurgical fired but did not cut
  - Point-of-care glucose results differ from lab results




## Examples of Medical Devices Problems



- **Interactions with other devices**
  - Burns with use of orthopedic shaver and grounding pad
  - Oxygen cylinder pulled inside MRI machine



## Examples of Medical Devices Problems



- **Use Errors :**

- Infusion pumps by the same manufacturer looks similar but operates differently
- Incorrectly positioned radiant warmer temperature sensor causes overheating




# Causes of Device Incident



# Causes of Device Incident

- Device related Factors
- External Factors
- Tampering and Sabotage
- Support System Failures
- User Factors



# Causes of Device Incident

## ☐ Device related Factors:

- Device failure
- labeling error
- Manufacturing error
- Packaging error
- Software deficiency
- Component failure
- Accessory Failure
- Device interactions
- Improper maintenance, testing, repair, or lack or failure of pre-use incoming inspection
- Improper modification

# Causes of Device Incident

## □ External Factors

- Power supply failure
- Medical gas/vacuum systems
- Electromagnetic EMI or Radio Frequency Interference RFI
- Environmental conditions
  - Temperature, humidity, light

# Causes of Device Incident

## Tampering and Sabotage (rare)

1. Family member
2. Patient
3. Healthcare workers (Doctor, Nurse, Aide)
4. Hacking (Cybersecurity)

# Causes of Device Incident

## ☐ **User Error** – 70% of device accidents

- Improper connection
- Incorrect clinical use
- Incorrect control settings
- Incorrect programming
- Failure to monitor
- Improper Maintenance

*Ref: 06 - Human Factors Considerations in Medical Device Accident Investigation - ECRI 2017.ppt*

# Causes of Device Incident

## Support System Failures


- Using inappropriate devices
- Improper storage
- Poor incident/recall reporting system
- Error in hospital policy

# Incident Reports

## Global and Local Statistics

# MD Incidents Reported in USA

Reports received by Report Type



<https://fis.fda.gov/sense/app/d10be6bb-494e-4cd2-82e4-0135608ddc13/sheet/7a47a261-d58b-4203-a8aa-6d3021737452/state/analysis>




# MD Incidents Reported in Australia

This figure illustrates the number of reports in The TGA's Incident Report and Investigation Scheme (IRIS) **by type of reporter** (2015-2017)

Source	2015	2016	2017
<b>Sponsors (AR)</b>	2817	3122	4604
<b>Doctors</b>	45	88	101
<b>Nurses</b>	134	157	109
<b>consumers</b>	88	97	205
<b>Allied health professionals</b>	266	347	351
<b>Total</b>	3359	3841	5370

<https://www.tga.gov.au/sites/default/files/medical-devices-post-market-vigilance-statistics-2016.pdf>

# Medical Devices Incidents Reports in KSA



# Incidents vs Adverse Events

# Exercises (Reportable or Not)

1. A **patient death** occurs following the use of external **defibrillator**. Device shocked appropriately but failed to bring back patient to sinus rhythm and patient expired.

 **NOT REPORTABLE**

2. A **patient death** occurs following the use of external **defibrillator**. A **malfunction** of devices contributed to the patient death.

 **REPORTABLE**

3. An infusion pump delivers the wrong dose because of an incompatibility between the pump and the infusion set used. The combination was made against the instructions for use. Medical intervention was required to avoid patient injury.

 **REPORTABLE**

## Exercises (Reportable or Not)

4. An infusion pump stop due to malfunction of the motor and fails to give an appropriate alarm.  
There is no patient injury.



5. The packaging of a sterile single use device is labeled with the caution: 'do not use if package is opened or damaged'.  
Prior to use, obvious damage to the packaging was observed and the device was not used.



6. The arm of an X-ray vascular system hand uncontrolled motion during patient examination.  
The patient was hit by the image intensifier and injured.




## Exercises (Reportable or Not)



8. A catheter fractured during insertion, with no suggestion of inappropriate handling. The fracture occurred in such a position that the broken part could easily be withdrawn. However, this was clearly a fortunate circumstance as if the catheter had fractured in a slightly different position then surgical intervention would have been necessary to retrieve the broken end.

# Exercises (Reportable or Not)

## 9. ISOLATION TRANSPORT BAG



REPORTABLE

NON-STERILE

# If AE occurred

## Immediate Actions Plan

- Protect patient and staff
- Protect equipment / environment
- Isolate equipment (including disposables)
- Record information
- Internal reporting (risk manager. Biomedical engineering)



# Preservation of Evidence

- Disposables - Save all!!
- Photographs – Take pictures
- Control Settings – Do not change
- Error codes in device memory!!!  
(event log) – Do not erase
- Isolating devices and Accessories



# Identify, What factors causing these AEs?



# Examples on AE Investigation

# Water Bag

## The result:

- Water bag has exploded on patient.
- A Second degree of burning on Back and Left arm.

## The Cause:

- Used High temperature water.
- Fill the bag 95% .
- Lie down top of a hot water bottle.
- There was no instruction for use with device.

## The action:

- Device Recall/correction.
- Add instruction for use.
- Safety Communication Notice.

الهيئة العامة للغذاء والدواء  
Saudi Food & Drug Authority

## كمادات الماء الساخن

### تعليمات عند شراء الكمامة

- التأكد من وجود تعليمات الاستخدام باللغة العربية.
- تتفالك كمادات الماء الساخن المصنوعة من مادة المطاط مع مرور الوقت بشكل أسرع، بينما تتدهور الكمامات المصنوعة من مادة البولي فينيل كلورايد (PVC) لفترة أطول.
- ينبغي اختيار الكمامات ذات الفتحات الكبيرة لأنها تخافس لتأثر الماء الساخن خلال ملئها.
- تأكد من خلوها من التسفقات أو العيوب المصنعية.

### ارشادات الاستخدام

- تعبئة الكمامة بعناية**  
تجنباً لتناثر أو انسكاب الماء الساخن بحيث لا يزيد مستوى الماء على ثلاثة أرباع حجمها.
- تفريغ الهواء من الكمامة**  
عن طريق خفصها بخذر حتى تصل إلى مستوى سطح مستو إلى أن يبدأ الماء بالخروج من الفتحة.
- اغلق الفتحة**  
بواسطة الغطاء (السدة) بإحكام.
- استخدم قطعة قماش**  
للإمسك بالكمامة بعددًا للترصبات أو الحروق.
- افتح الغطاء (السدة)**
- احرص على إفراغ الكمامة**  
من المياه (عند تكرار الاستخدام).
- استخدم كبار السن والأطفال**  
للكمادات يكون بإشراف شخص بالغ لتجنب الاستخدام الخاطئ.

### تحذيرات عند الاستخدام

- التحقق من جودة الكمامة في حال استخدامها.
- عدم استخدام الماء المغلي بالكمامة لأنه قد يسبب إصابات وحروقاً بالجلد ويقلل من عمرها الافتراضي.
- لا تجلس أو تتمد فوق الكمامات، لأنها لا تتحمل الضغط.

بالأهم نتمتع  
#الغذاء\_والدواء

Saudi\_FDA  
www.sfda.gov.sa

SFDA 19999  
مركز تقييم الجودة

# Transport incubator

## The result:

- A burn happened with transport incubator while Transport patient from hospital to another hospital.

## The Cause:

- The unit is 120V .
- One of the staff nurses changed the power cable from US to UK and plugged it into 220V socket.

## The action:

Developing written procedures for handling medical devices



# Patient Monitor

## The result:

- Suddenly freeze and the image could turn black and there is no more visual representation of the patient data on this monitor.
- Nevertheless, the Patient Monitoring system keeps on working and the audible alarms remain.

## The cause:

- electrolytic capacitor on the **inverter board**

## The action:

- SFDA have Request to Manufacturer to Submit a FSN and taking Corrective action.



## The result:

- Death report
- Investigation team carried out investigations that included checking the device, checking event data stored in the machine's log, and gathering information about the incident.

## The cause:

- From the error log , alarm On AC power failure and alarm Off AC power failure are showed many times.
- The device was working on the battery mode.
- Medical staff did not respond to the device alarm and the device stopped working for ten minutes.
- There was a malfunction in the electrical socket

## Action:

- Change the electrical socket with new one

# Ventilator

```

04-Aug-2013,23:59:20,PMB Alarm On: AC power failure
04-Aug-2013,23:59:21,On Battery
05-Aug-2013,00:00:00,PMB Alarm Off: AC power failure
05-Aug-2013,00:00:17,02 Supply Pres Low
05-Aug-2013,00:00:17,Total Flow Sensor Communications Failure
05-Aug-2013,00:10:34,PMB Alarm On: AC power failure
05-Aug-2013,00:10:35,On Battery
05-Aug-2013,01:17:04,Low Internal Battery - 30 Min
05-Aug-2013,01:17:35,On Battery
05-Aug-2013,01:18:35,Low Internal Battery - 30 Min
05-Aug-2013,01:19:34,On Battery
05-Aug-2013,01:20:04,Low Internal Battery - 30 Min
05-Aug-2013,01:25:35,Low Internal Battery - 20 Min
05-Aug-2013,01:26:35,Low Internal Battery - 30 Min
05-Aug-2013,01:27:35,Low Internal Battery - 20 Min
05-Aug-2013,01:32:35,Low Internal Battery - 10 Min
05-Aug-2013,01:34:05,Low Internal Battery - 5 Min
05-Aug-2013,01:34:33,PMB Alarm On: 1 min time left
05-Aug-2013,01:34:34,Low Internal Battery - 1 Min
05-Aug-2013,01:53:18,*** Main App Started ***
05-Aug-2013,01:54:39,PMB Saved Alarm Off: AC power failure
05-Aug-2013,02:17:54,Standby Patient Detection
05-Aug-2013,02:17:54,Circ Unparked Patient Detect
05-Aug-2013,04:06:26,*** Main App Started ***
05-Aug-2013,04:06:57,Standby Patient Detection
05-Aug-2013,04:08:14,Therapy Power Off
05-Aug-2013,04:08:37,Standby Patient Detection
05-Aug-2013,04:08:37,Circ Unparked Patient Detect
05-Aug-2013,04:10:22,Standby Patient Detection
05-Aug-2013,04:10:22,Circ Unparked Patient Detect
05-Aug-2013,04:10:31,Standby Patient Detection
05-Aug-2013,04:10:31,Circ Unparked Patient Detect
05-Aug-2013,05:13:13,PMB Alarm On: AC power failure
05-Aug-2013,05:13:14,On Battery
05-Aug-2013,05:22:12,Low Internal Battery - 30 Min
05-Aug-2013,05:23:43,On Battery
    
```

the ventilator was in the battery mode on 05 August 2013, at 12:10 AM till the ventilator shutdown at 1:43 AM and connected to the outlet at 1:54 AM



# Anesthesia

## The result:


- Death report.

## The cause:

- The device works properly and the problem is caused by the user forgetting to reconnect the fresh air hose to the external fresh gas outlet after the previous case.
- Lack of sufficient fresh air reduces the concentration of oxygen in the inhalation, leading to insufficient ventilation for the patient.

## The action:

- User training







# Risk Classification of Adverse Events

# Risk Classification of Adverse Events

Calculating risk score using special formula (below). There are four factors influence the risk score:

- A-Severity of injury,
- B-Detectability of occurrence,
- C-Type of incident / likelihood of occurrence
- D-Impact on the Saudi market.

$$\text{SFDA risk Score} = (A+B+C) * D$$



# Risk Classification of Adverse Events

By calculating the SFDA risk score we can determine the risk level

Risk Score	Risk level
3-19	Minor incident (Level1)
20-34	Serious incident (Level 2)
38-48	National crisis (Level3)



# Determining the Severity of injury (A)



## 1. Limited

- No injury
- Temporary
- Self-limiting
- Minor injury could occur

## 2. Moderate

- Prolonged hospitalization
- Medical intervention required
- Delay in diagnosis or treatment
- Incorrect diagnosis or treatment
- Packaging failure or potential sterility issue
- Known severe sterility issue

## 3. Severe

- Permanent injury and/or impairment
- Revision of implantable device
- Surgical intervention required
- Delay in critical diagnosis, therapy or treatment
- Incorrect critical diagnosis or treatment

## 4. Life threatening

- Life threatening injury
- Transmission of life threatening infection
- Death

# Determining the Detectability of occurrence (B)

## 1. Very likely

- Fault with the device is very obvious
- The device cannot be used

## 2. Likely

- Fault with the device is visible

## 3. Unlikely

- It is not possible to see or detect the fault with the device before use

## 4. Extremely unlikely

- It is not possible to see or detect the fault with the device at all



# Determining the Type of Action / likelihood of occurrence(C)

## **1. Trend reporting /User training**

- No similar occurrences
- Occurrence appears isolated
- User error suspected
- Abnormal use

## **2. Low risk FSCA**

- Limited number of similar occurrences
- HCP can intervene / prevent adverse outcome

## **3. Medium risk FSCA**

- High number of similar occurrences
- Single severe occurrence
- HCP may not be able to intervene

## **4.High risk FSCA**

- Multiple severe occurrences
- Occurrence is not obvious to HCP
- Intervention is not be possible to prevent serious adverse outcome

# Determining the Impact on Saudi market (D)

## **1. Limited impact**

- Manufacturer trend reports
- Signal reports

## **2. Low impact**

- Low risk FSCA affecting Saudi market
- Low risk Saudi incident
- Local incident directly linked to ongoing FSCA / PSR
- Saudi leads CA (Saudi market not affected)

## **3. Moderate impact**

Medium risk FSCA affecting Saudi market

Medium risk Saudi incident

## **4. High impact**

- High risk FSCA affecting Saudi market
- High risk Saudi incident
- FSCA affecting Saudi market with associated Saudi incident(s)
- Saudi leads CA (Saudi market affected)
- Highly involved issue
- Serious public health issue
- Death directly linked to device



## Action taken after calculation:-

After calculating the risk score, there are different action depend on the risk levels as follow:

**1-** All Minor incidents (Level1) considered as normal tasks “routine”.

**2-** For Serious incident (Level 2) or National crisis (Level3) will activate the crisis management plan “emergency plan”; and contact Crisis Department.

**3-** In case of adverse events that meet the below conditions, it will transferred to Biometrics department for further evaluation:

- Adverse Events that are frequently occurring.
- Adverse events that have potential public health impact.
- some cases that the assessment’s team found there is a need to do more clinical trials to ensure safety.

# FSN's/ FSCAs Management

# What is FSN?

## Field Safety Notice (FSN)

A communication to customers and/or users sent out by a manufacturer or its representative in relation to a Field Safety Corrective Action (FSCA).

# What is a Field Safety Corrective Action ?

## Field Safety Corrective Action

An action taken by a manufacturer to reduce or remove a risk of death or serious deterioration in the state of health associated with the use of a medical device that is ALREADY placed on the market. Such actions are notified via a Field Safety Notice.

## What type of actions is required?

- **Correction:**

Addresses a problem with a medical device in the place where it is used or sold.

- **Removal:**

Addresses a problem with a medical device by removing it from where it is used or sold.


# Case studies

## Case Study (from AE to FSN)

**Affected Product:** Ultrasound Probe Cover Kit with Gel Manufactured by YAFHO BIO-TECHNOLOGY CO.

### Investigation Summary:

- Start investigation on 8 Jun 2016 and investigator visited the hospital on 9 Jun 2016
- Tests Performed: Sterility test and Organism identification test
- Result showed: Positive contamination of (*Burkholderia cepacia*) in two lot number.
- Inform POE to stop permeation of device to enter in Saudi market on 14 Jun 2016
- Publish official letter to health care provider officer on 14 Jun 2016.
- Publish safety communication on SFDA web Site on 26 Jun 2016.
- Manufacturer issued a recall about all affected lots



ULTRASOUND PROBE COVER TELESCOPIC FOLDED WITH GEL من طراز 654005، وهو غطاء بلاستيكي لمجس جهاز الموجات فوق الصوتية	اسم المنتج
20160201 / 20150701	أرقام الترخيلات
YAFHO BIO-TECHNOLOGY CO, INC - الصين	اسم المصنع
وجود بكتيريا من نوع <i>Burkholderia Cepacia</i>	سبب التحذير

### الإبلاغ عن المشاكل المتعلقة بالأجهزة والمنتجات الطبية

المركز الوطني لبلاغات الأجهزة والمنتجات الطبية  
البريد الإلكتروني: <http://ncmdr.sfda.gov.sa>  
مركز الاتصال الموحد 19999

# Case Study (Human Factors / Technique of Use)

– Defibrillator Allegedly Malfunctioned during Resuscitation

## Brief Overview:

- Patient required synchronized cardioversion
- First sync shock successful
- Second sync shock attempt – defibrillation shock delivered!
- Patient went into ventricular fibrillation
- Patient died
- **Hospital tested the defibrillator and found it to be functioning properly**

## Defibrillator Information given by Hospital:

- Sync markers were “temporarily” displayed on the screen
- Electrodes were used (not paddles)
- 1st shock attempt – 200 J, 2nd attempt – 300 J.



Ref:22-2 - Defibrillator - Hypothetical Scenarios – ECRI 2017



## Case Study (Human Factors / Technique of Use)

### Design Considerations



- Same manufacturer
- Different design

# Safety Communications

# Safety Communication

- **Safety communication:** a communication provides an important information and recommendations about safety of medical devices, issued by Medical Devices Sector within Saudi-FDA and addressed to public, healthcare providers and/ or health professional.
- These notifications does not necessarily mean a medical device/product is considered to be unsafe
- HealthCare Professional or Public

# Why Reporting is Important

- **Patient safety**  
Prevent future problems and protect our patients, staff, families, and visitors
- **Provide information to manufacturers and/or Saudi Food and Drug Authority**(Build database, monitoring trends)
- **Device Regulation**  
Effect changes in policies and procedures
- **Accrediting agencies** Ex: CBAHI
- **Assist Risk Management with claims or litigation**

# Proactive Activities

# Definitions

## Clinical Investigation

Any systematic investigation or study in or on one or more human subjects, undertaken to assess the safety or performance of a medical device. .

## Clinical Evaluation

An assessment and analysis of clinical data pertaining to a medical device in order to verify the clinical safety and performance of the device

## Clinical Data

The safety and performance information that is generated from the use of the medical device.

### Source of the clinical data:

- Clinical Investigation of the device
- Literature review
- Published or unpublished reports on other clinical experience of either the device or similar device

**Which medical devices require a clinical evaluation?**


**All medical devices regardless of their  
classification**




## When is the clinical evaluation undertaken?

- An ongoing process, throughout the device's life cycle
- During the conformity assessment process ( prior to market the device)
- Mandatory to get CE and must be updated afterwards
- Insufficient clinical evidence. ( Approval by equivalent ...)


## What should the clinical evaluator consider (address) ??




# Clinical Evaluation and Investigation routes




# Clinical Evaluation stages



## Relevant Clinical Data



## Consultation and Feedback



## Analysis of the clinical data

The goal of the analysis stage is to determine if the collected clinical data answer the unanswered question regarding the safety and effectiveness of the device

## Recommendation and conclusion

- Gather all the deliverable outcomes from the clinical evaluation in the discussion and recommendation section
- Summarize the required action based on the analyzed data in the conclusion section



# Safe use of Medical Devices within Healthcare Facilities



# Monitoring the Performance of Medical Devices within hospitals

	Current	Future
<p><b>Monitoring the Performance of Medical Devices within hospitals</b></p>	<ul style="list-style-type: none"> <li>No National requirements for safe use of medical devices within the hospitals in KSA.</li> <li>There is no effective evaluation process for performance of medical devices during their operation.</li> </ul>	<ul style="list-style-type: none"> <li>Promote the quality and efficiency of health care through implementing the SFDA Requirements for Quality, safety and effectiveness of medical devices at hospitals</li> <li>Contribute to reduce the incidents related to calibration, maintenance, poor storage, transportation and disposal medical devices.</li> </ul>



## Why are we evaluating the use of MDs at healthcare facilities? (Objectives)

- To develop unified national requirements for best practices in medical device management
- To increase the lifetime and reduce the overall cost of health technologies within healthcare facilities.
- To mitigate the risks associated with the usage, transportation and storage of medical devices within healthcare facilities.
- To ensure that medical devices within healthcare facilities:
  - Are utilized appropriately and effectively;
  - Are maintained in a safe and reliable condition;
  - Are operated in accordance with the manufacturer's instructions by trained users and professionals
- To facilitate the periodic monitoring medical devices in relation to the following:
  - Performance and safety inspections;
  - Calibration
  - Reporting and tracking of related adverse events.

# Advantages of following the requirements

- **Safety**
  - patients
  - operators
  - others
- **efficiency**
  - Enhance the quality of care.
- **Economic**
  - Extend the life of the device
  - proper management of property
  - Unifying the criteria and procedures of safety and performance.

## How are we performing the evaluation?

- Through,
  - Developing a guideline that specifies the SFDA requirements,
  - Delivering the requirements into healthcare facility in a first meeting, and
  - Evaluating the implementation of such requirements in a follow-up evaluation

# How are we performing the evaluation?

## Policies, Procedures, and implementation in:

- Establishments & Products Approvals
- AEs Reporting
- Follow-up with SFDA Recalls
- Clinical investigation & Ads. Approvals

The advertisement for Anuice hemorrhoid treatment includes the following text:
   
"أحدث جهاز أمريكي لعلاج البواسير بدون جراحة" (New American device for hemorrhoid treatment without surgery)
   
"أنيويس" (Anuice)
   
"تنتائج فورية تبدأ من اليوم الأول" (Immediate results start from the first day)
   
"متوفر بالصيدليات" (Available in pharmacies)
   
"يو صبي به الأظباء" (Dr. Youssef bin al-Azba)
   
"لا يوجد لدينا وكلاء أو موزعين داخل المملكة العربية السعودية" (We do not have agents or distributors in Saudi Arabia)
   
"www.anuice.com.sa"
   
"MEDASSIST"
   
"MDNR No Approval (MDNR)"
   
"Treatment for Hemorrhoids without the effects of surgery"
   
Annotations:
   
- A red circle highlights "تنتائج فورية تبدأ من اليوم الأول" with the text: "The applicant Claimed 'Fast' result since the first day BUT there is NO evidence from manufacturer"
   
- A red circle highlights "MDNR No Approval (MDNR)" with the text: "The MDNR shall not be printed on advertising material"



# How are we performing the evaluation?

## Policies, Procedures, and implementation in:

- Asset Management
- BME/BMT Qualifications & Trainings
- Manuals Availability
- Well-Equipped Workshop(s)
- Electrical Safety Tests
- Intended and Proper Use
- Appropriate Spare Parts
- Recommended accessories
- PPM and CM
- Calibration Tests
- Transport & Storage
- Disposal



### PREVENTIVE MAINTENANCE

I.D. \_\_\_\_\_

BY \_\_\_\_\_ DATE \_\_\_\_\_

DUE \_\_\_\_\_

# What are the major outcomes (so far)?

- Most frequent observations:
  - Electricity (plug & socket compatibility)
  - Missing PPM tags
  - Lack of user and service manuals
  - Lack of user and service training





## What are the major outcomes (so far)?

- Most frequent observations:
  - Expired oxygen sensors (ventilators)



# Postmarket Surveillance



# Thank You

**ANY  
QUESTIONS?**

